

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4116

Name: RITTER2  
Sterilizer ID: V2313228  
11/13/2024 05:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.7 F 0.0  
2:00 157.6 F 1.5  
4:00 168.7 F 5.0  
6:00 196.1 F 9.0  
8:00 211.6 F 13.3  
10:00 221.6 F 7.5  
12:00 232.9 F 12.3  
14:00 235.5 F 9.6  
16:00 238.1 F 9.6  
18:00 247.8 F 14.5  
20:00 259.3 F 20.9  
22:00 269.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 272.9 F 30.1  
2:00 271.7 F 29.5  
3:00 273.0 F 30.2  
4:00 272.1 F 29.5

Min 271.6 F 29.4  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 217.0 F 0.0  
10:00 228.1 F 0.0  
15:00 230.1 F 0.0  
20:00 234.2 F 0.0  
25:00 226.9 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 23:47  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:02:45  
CYCLE COMPLETE

11/13/2024 06:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_