
Midmark M11 - v1.0.0
Total Cycles: 4251

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2024 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	183.9 F	0.0
2:00	173.8 F	1.3
4:00	187.8 F	4.6
6:00	201.1 F	8.8
8:00	215.7 F	13.4
10:00	226.3 F	6.5
12:00	236.7 F	10.8
14:00	231.8 F	7.4
16:00	244.3 F	12.6
18:00	241.7 F	11.0
20:00	253.5 F	16.9
22:00	263.8 F	23.0
24:00	272.3 F	28.3

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.1
1:00	272.4 F	28.5
2:00	271.4 F	28.4
3:00	272.4 F	28.4
4:00	271.2 F	28.4

Min 271.0 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	196.6 F	0.0
	5:00	214.9 F	0.0
	10:00	237.4 F	0.0
	15:00	233.8 F	0.0
	20:00	228.3 F	0.0
	25:00	206.1 F	0.0
Drying Complete	30:00	200.8 F	0.0

FILLING: 1:26
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

01/31/2024 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____