
Midmark M11 - v1.0.5
Total Cycles: 3918

Name: RITTER2
Sterilizer ID: V2313228
09/09/2024 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.1 F 0.0
2:00 160.7 F 2.3
4:00 165.7 F 5.7
6:00 189.0 F 10.0
8:00 214.8 F 6.7
10:00 225.1 F 9.2
12:00 233.0 F 10.2
14:00 239.2 F 12.1
16:00 239.7 F 10.9
18:00 252.4 F 17.5
20:00 263.7 F 24.3
22:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.0
1:00 272.1 F 29.9
2:00 272.0 F 29.9
3:00 272.5 F 30.1
4:00 272.3 F 30.4

Min 271.7 F 29.9
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 223.0 F 0.0
10:00 232.9 F 0.0
15:00 227.7 F 0.0
20:00 230.2 F 0.0
25:00 224.5 F 0.0
30:00 218.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:44
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

09/09/2024 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____