
Midmark M11 - v1.0.5
Total Cycles: 1182

Name: RITTER2
Sterilizer ID: V2313228
06/09/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 176.9 F 3.5
4:00 192.9 F 8.5
6:00 212.7 F 9.2
8:00 227.6 F 9.6
10:00 234.1 F 9.1
12:00 242.7 F 12.6
14:00 244.6 F 13.2
16:00 257.7 F 20.2
18:00 269.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.1 F 29.9
2:00 272.7 F 29.8
3:00 272.3 F 29.4
4:00 271.7 F 29.0

Min 271.1 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 226.1 F 0.0
10:00 222.9 F 0.0
15:00 220.5 F 0.0
20:00 213.5 F 0.0
25:00 209.1 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

06/09/2022 12:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____