
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003117
15:20 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:32
STERILIZING 3:32
136.4 C 223kPa 3:32
136.3 C 225kPa 4:02
136.3 C 219kPa 4:32
136.4 C 224kPa 5:02
136.3 C 222kPa 5:32
136.4 C 222kPa 6:02
136.4 C 222kPa 6:32
136.4 C 224kPa 7:02
136.4 C 224kPa 7:02
VENTING 7:02
AIR DRYING 7:56
CYCLE COMPLETE 8:55

Digital Signature #
31907572AFBCBF3A
