
Midmark M11 - v1.0.5
Total Cycles: 232

Name: RITTER2
Sterilizer ID: V2313228
07/21/2021 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 149.9 F 2.0
4:00 155.0 F 5.4
6:00 179.6 F 9.7
8:00 203.1 F 7.2
10:00 215.2 F 8.6
12:00 230.2 F 12.2
14:00 232.2 F 11.3
16:00 237.5 F 9.9
18:00 251.3 F 16.3
20:00 260.4 F 22.8
22:00 271.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.9 F 30.4
1:00 272.8 F 30.5
2:00 272.4 F 30.4
3:00 271.7 F 30.3
4:00 271.8 F 30.7

Min 271.4 F 30.3
Max 273.1 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 207.4 F 0.0
10:00 228.6 F 0.0
15:00 227.5 F 0.0
20:00 222.4 F 0.0
25:00 219.7 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

07/21/2021 01:19 PM
ACCEPTED: X
REJECTED:
Operator: U01