
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002513
15:35 11/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:15
136.5 C 225kPa 4:15
136.2 C 224kPa 4:45
136.4 C 225kPa 5:15
136.3 C 224kPa 5:45
136.4 C 226kPa 6:15
136.4 C 226kPa 6:45
136.4 C 226kPa 7:15
136.3 C 224kPa 7:45
136.3 C 224kPa 7:45
VENTING 7:45
AIR DRYING 8:40
CYCLE COMPLETE 8:50

Digital Signature #
9BCCE0066B0DF8F9
