

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3677

Name: RITTER 1  
Sterilizer ID: V2178198  
08/18/2023 02:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.3 F 0.0  
2:00 150.5 F 2.2  
4:00 157.6 F 5.2  
6:00 184.9 F 8.6  
8:00 199.6 F 13.0  
10:00 216.0 F 6.3  
12:00 230.7 F 10.6  
14:00 231.7 F 7.4  
16:00 240.7 F 12.1  
18:00 239.0 F 10.2  
20:00 251.5 F 16.2  
22:00 261.8 F 22.2  
24:00 271.4 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.5 F 29.3  
2:00 272.3 F 29.6  
3:00 271.5 F 28.9  
4:00 272.3 F 29.3

Min 271.5 F 28.9  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 214.7 F 0.0  
10:00 221.5 F 0.0  
15:00 214.7 F 0.0  
20:00 211.7 F 0.0  
25:00 208.8 F 0.0  
30:00 205.8 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:14  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:03:55  
CYCLE COMPLETE

08/18/2023 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_