
Midmark M11 - v1.0.0
Total Cycles: 5144

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2024 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.4 F 0.0
2:00 123.5 F 0.0
4:00 147.6 F 0.2
6:00 182.0 F 0.8
8:00 204.6 F 4.1
10:00 222.3 F 8.4
12:00 231.6 F 13.4
14:00 235.7 F 9.9
16:00 235.5 F 8.3
18:00 241.6 F 10.8
20:00 247.3 F 13.9
22:00 260.3 F 21.0
24:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.1 F 28.9
4:00 272.4 F 28.9

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 214.9 F 0.0
10:00 214.5 F 0.0
15:00 203.7 F 0.0
20:00 219.1 F 0.0
25:00 220.5 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:56
CYCLE COMPLETE

11/29/2024 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____