

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2037

Name: RITTER2  
Sterilizer ID: V2313228  
02/09/2023 07:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
2:00 202.5 F 4.5  
4:00 214.6 F 11.0  
6:00 227.7 F 7.7  
8:00 239.1 F 11.1  
10:00 245.3 F 13.1  
12:00 246.8 F 13.8  
14:00 261.1 F 21.8  
16:00 272.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.6 F 29.1  
2:00 272.8 F 29.7  
3:00 272.0 F 28.8  
4:00 273.0 F 29.6  
  
Min 271.7 F 28.7  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 211.5 F 0.0  
10:00 227.3 F 0.0  
15:00 228.0 F 0.0  
20:00 224.8 F 0.0  
25:00 222.6 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 2:45  
HEATING: 16:16  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:56:03  
CYCLE COMPLETE

02/09/2023 08:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_