
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006413
10:49 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 3:57
136.0 C 220kPa 3:57
136.2 C 222kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
136.1 C 222kPa 5:57
136.4 C 224kPa 6:27
136.4 C 224kPa 6:57
136.3 C 222kPa 7:27
136.3 C 222kPa 7:27
VENTING 7:27
AIR DRYING 8:20
CYCLE COMPLETE 25:00

Digital Signature #
905BEFB1B3FF0578
