
Midmark M11 - v1.0.0
Total Cycles: 1854

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2022 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 173.3 F 3.5
4:00 197.7 F 8.4
6:00 215.6 F 10.0
8:00 229.2 F 8.1
10:00 240.7 F 13.2
12:00 239.4 F 10.4
14:00 236.1 F 8.4
16:00 250.7 F 15.4
18:00 262.6 F 22.1
20:00 272.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.8
2:00 271.6 F 27.9
3:00 272.7 F 28.9
4:00 272.3 F 28.4

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 218.6 F 0.0
10:00 241.8 F 0.0
15:00 227.3 F 0.0
20:00 218.3 F 0.0
25:00 195.1 F 0.0
Dry Cycle Aborted

FILLING: 1:56
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:13
DRYING: 25:18
TOTAL CYCLE: 00:54:44

04/21/2022 11:58 AM
ACCEPTED: X
REJECTED:
Operator: U02