
Midmark M11 - v1.0.5
Total Cycles: 87

Name: RITTER2
Sterilizer ID: V2313228
06/10/2021 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 154.2 F 1.8
4:00 158.7 F 4.6
6:00 181.2 F 8.4
8:00 198.0 F 12.8
10:00 221.9 F 6.5
12:00 231.0 F 11.2
14:00 233.8 F 8.1
16:00 245.4 F 13.2
18:00 244.3 F 12.4
20:00 255.8 F 18.4
22:00 266.1 F 24.8
24:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 271.5 F 29.0
2:00 271.5 F 29.1
3:00 271.5 F 29.0
4:00 271.5 F 28.9

Min 271.4 F 28.8
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 231.0 F 0.0
10:00 225.0 F 0.0
15:00 227.9 F 0.0
20:00 224.3 F 0.0
25:00 215.5 F 0.0
30:00 211.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

06/10/2021 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____