
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003167
11:09 16/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:00
STERILIZING 4:04
136.5 C 220kPa 4:04
136.3 C 223kPa 4:34
136.2 C 219kPa 5:04
136.4 C 224kPa 5:34
136.2 C 222kPa 6:04
136.4 C 224kPa 6:34
136.4 C 223kPa 7:04
136.3 C 218kPa 7:34
136.3 C 218kPa 7:34
VENTING 7:35
AIR DRYING 8:32
CYCLE COMPLETE 8:35

Digital Signature #
717F4AEB16C23D40
