

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4088

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2023 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 72.4 F 0.2  
4:00 83.5 F 1.2  
6:00 116.1 F 3.2  
8:00 143.5 F 5.6  
10:00 163.4 F 8.4  
12:00 193.0 F 11.2  
14:00 205.1 F 3.9  
16:00 209.4 F 4.7  
18:00 218.4 F 7.8  
20:00 227.9 F 11.3  
22:00 228.2 F 6.5  
24:00 238.8 F 11.1  
26:00 234.5 F 8.1  
28:00 247.7 F 13.9  
30:00 257.8 F 19.4  
32:00 267.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.4 F 29.0  
2:00 272.2 F 28.6  
3:00 271.1 F 28.0  
4:00 271.6 F 28.6  
  
Min 270.9 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 220.3 F 0.0  
10:00 225.1 F 0.0  
15:00 229.1 F 0.0  
20:00 230.0 F 0.0  
25:00 231.7 F 0.0  
30:00 228.8 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:38  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:12:09  
CYCLE COMPLETE

12/06/2023 10:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_