
Midmark M11 - v1.0.5
Total Cycles: 2979

Name: RITTER2
Sterilizer ID: V2313228
10/05/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 180.2 F 3.4
4:00 201.0 F 8.3
6:00 217.8 F 13.3
8:00 227.8 F 8.5
10:00 236.2 F 11.5
12:00 241.3 F 12.3
14:00 241.7 F 11.6
16:00 254.3 F 18.2
18:00 265.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 271.8 F 29.2
2:00 272.1 F 29.5
3:00 272.3 F 29.9
4:00 271.5 F 29.4

Min 271.4 F 29.2
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 218.9 F 0.0
10:00 222.0 F 0.0
15:00 214.7 F 0.0
20:00 208.8 F 0.0
25:00 206.2 F 0.0
30:00 203.7 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

10/05/2023 01:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____