
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001538
8:29 22/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:02
STERILIZING 4:11
136.3 C 222kPa 4:11
136.4 C 222kPa 4:41
136.2 C 221kPa 5:11
136.4 C 222kPa 5:41
136.4 C 224kPa 6:11
136.4 C 223kPa 6:41
136.4 C 224kPa 7:11
136.1 C 219kPa 7:41
136.1 C 219kPa 7:41
VENTING 7:41
AIR DRYING 8:35
CYCLE COMPLETE 11:12

Digital Signature #
AA23A5AECA1F3D47
