
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000890
10:43 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 4:06
STERILIZING 5:37
136.0 C 221kPa 5:37
136.2 C 229kPa 6:07
136.4 C 232kPa 6:37
136.2 C 230kPa 7:07
136.1 C 230kPa 7:37
136.4 C 232kPa 8:07
136.1 C 229kPa 8:37
136.2 C 230kPa 9:07
136.2 C 230kPa 9:07
VENTING 9:07
AIR DRYING 10:13
CYCLE COMPLETE 13:57

Digital Signature #
D5C18D5B05C3161C
