

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1187

Name: RITTER2  
Sterilizer ID: V2313228  
06/15/2022 07:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 74.7 F 0.0  
4:00 105.7 F 0.0  
6:00 121.6 F 0.9  
8:00 147.2 F 4.0  
10:00 170.0 F 7.5  
12:00 190.0 F 11.4  
14:00 207.1 F 3.5  
16:00 219.6 F 7.1  
18:00 230.6 F 10.9  
20:00 231.7 F 7.6  
22:00 242.0 F 12.4  
24:00 240.3 F 10.7  
26:00 252.9 F 16.9  
28:00 263.7 F 23.4  
30:00 271.8 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.2 F 28.9  
2:00 272.5 F 29.4  
3:00 271.5 F 28.6  
4:00 272.3 F 29.0

Min 271.4 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
5:00 208.9 F 0.0  
10:00 207.7 F 0.0  
15:00 208.3 F 0.0  
20:00 210.9 F 0.0  
25:00 213.2 F 0.0  
30:00 213.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:08  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 01:07:57  
CYCLE COMPLETE

06/15/2022 08:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_