

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3782

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2024 01:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.0 F 0.0  
2:00 150.8 F 2.2  
4:00 157.3 F 5.5  
6:00 182.3 F 9.9  
8:00 209.9 F 6.6  
10:00 218.9 F 9.1  
12:00 235.5 F 10.4  
14:00 237.1 F 11.7  
16:00 238.0 F 10.5  
18:00 251.5 F 17.1  
20:00 263.0 F 23.8  
22:00 272.0 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 272.0 F 29.7  
2:00 272.3 F 30.1  
3:00 271.2 F 29.5  
4:00 272.1 F 29.9

Min 271.2 F 29.4  
Max 272.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 213.3 F 0.0  
10:00 221.2 F 0.0  
15:00 219.9 F 0.0  
20:00 226.8 F 0.0  
25:00 228.8 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 22:09  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:00:58  
CYCLE COMPLETE

07/15/2024 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_