

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3322

Name: RITTER2  
Sterilizer ID: V2313228  
01/25/2024 11:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.4 F 0.0  
2:00 151.9 F 1.1  
4:00 154.6 F 4.0  
6:00 176.8 F 7.9  
8:00 198.7 F 12.7  
10:00 215.7 F 6.4  
12:00 230.2 F 11.4  
14:00 230.7 F 8.0  
16:00 241.7 F 12.0  
18:00 244.1 F 12.9  
20:00 256.7 F 19.5  
22:00 268.0 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.5 F 29.5  
2:00 271.4 F 29.1  
3:00 272.3 F 29.5  
4:00 271.3 F 28.9

Min 271.3 F 28.9  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 205.1 F 0.0  
10:00 214.9 F 0.0  
15:00 206.3 F 0.0  
20:00 199.1 F 0.0  
25:00 196.0 F 0.0  
30:00 193.9 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:17  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

01/25/2024 12:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_