

-----  
Midmark M11 - v1.0.0  
Total Cycles: 992

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2021 07:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.8 F 0.0  
2:00 169.1 F 1.5  
4:00 184.6 F 4.6  
6:00 196.9 F 9.0  
8:00 218.9 F 8.7  
10:00 226.9 F 7.8  
12:00 237.3 F 12.9  
14:00 238.1 F 9.7  
16:00 237.6 F 8.8  
18:00 247.8 F 13.9  
20:00 259.4 F 20.2  
22:00 269.8 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.7 F 28.7  
2:00 272.5 F 29.2  
3:00 271.7 F 28.4  
4:00 272.7 F 29.0

Min 271.5 F 28.4  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
5:00 234.4 F 0.0  
10:00 215.9 F 0.0  
15:00 209.0 F 0.0  
20:00 201.1 F 0.0  
25:00 210.6 F 0.0  
Dry Cycle Aborted

FILLING: 0:54  
HEATING: 23:40  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 27:26  
TOTAL CYCLE: 00:59:06

07/12/2021 08:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01