
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000832
15:19 22/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:06
STERILIZING 4:15
136.0 C 228kPa 4:15
136.2 C 230kPa 4:45
136.3 C 232kPa 5:15
136.4 C 233kPa 5:45
136.4 C 229kPa 6:15
136.2 C 230kPa 6:45
136.4 C 232kPa 7:15
136.4 C 232kPa 7:45
136.4 C 232kPa 7:45
VENTING 7:45
AIR DRYING 8:38
CYCLE COMPLETE 19:52

Digital Signature #
D9936A930D89118F
