

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2397

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2022 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.8 F 0.0  
2:00 154.6 F 2.4  
4:00 173.3 F 5.9  
6:00 196.0 F 10.4  
8:00 216.1 F 5.1  
10:00 229.3 F 9.8  
12:00 232.4 F 8.3  
14:00 242.3 F 12.5  
16:00 241.2 F 10.9  
18:00 254.6 F 17.7  
20:00 266.2 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.3 F 29.0  
2:00 271.4 F 28.3  
3:00 272.5 F 29.2  
4:00 271.7 F 28.4  
  
Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 229.3 F 0.0  
10:00 234.8 F 0.0  
15:00 229.5 F 0.0  
20:00 226.6 F 0.0  
25:00 227.3 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:35  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:51  
CYCLE COMPLETE

10/04/2022 02:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_