
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000056
14:31 19/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:11
STERILIZING 4:14
136.0 C 220kPa 4:14
136.4 C 222kPa 4:44
136.0 C 222kPa 5:14
136.4 C 224kPa 5:44
136.0 C 223kPa 6:14
136.4 C 225kPa 6:44
136.3 C 224kPa 7:14
136.3 C 224kPa 7:44
136.3 C 224kPa 7:44
VENTING 7:44
AIR DRYING 8:38
CYCLE COMPLETE 11:27

Digital Signature #
9C1FD0911971B377
