
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001854
14:41 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:18
STERILIZING 3:30
136.4 C 227kPa 3:30
136.4 C 227kPa 4:00
136.1 C 224kPa 4:30
136.3 C 225kPa 5:00
136.4 C 226kPa 5:30
136.4 C 227kPa 6:00
136.2 C 223kPa 6:30
136.4 C 227kPa 7:00
136.4 C 227kPa 7:00
VENTING 7:00
AIR DRYING 8:00
CYCLE COMPLETE 8:03

Digital Signature #
3052E70DE281EB03
