
Midmark M11 - v1.0.5
Total Cycles: 3117

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 177.8 F 3.1
4:00 196.8 F 8.3
6:00 215.8 F 11.5
8:00 224.0 F 9.2
10:00 233.8 F 9.6
12:00 242.4 F 13.0
14:00 243.2 F 12.4
16:00 256.6 F 19.5
18:00 268.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.2 F 29.9
2:00 271.9 F 29.2
3:00 271.8 F 29.7
4:00 272.4 F 29.6

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 232.5 F 0.0
10:00 223.0 F 0.0
15:00 219.7 F 0.0
20:00 221.2 F 0.0
25:00 217.8 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

11/13/2023 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____