

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1026

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2022 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.5 F 0.0  
2:00 155.2 F 2.6  
4:00 164.4 F 6.4  
6:00 199.8 F 11.6  
8:00 214.2 F 6.1  
10:00 229.1 F 11.3  
12:00 231.7 F 8.5  
14:00 241.0 F 11.0  
16:00 245.2 F 13.6  
18:00 257.7 F 20.4  
20:00 268.8 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.1 F 30.1  
2:00 272.8 F 30.0  
3:00 272.2 F 29.6  
4:00 271.4 F 29.2  
  
Min 271.2 F 29.1  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 206.7 F 0.0  
10:00 217.5 F 0.0  
15:00 223.6 F 0.0  
20:00 225.4 F 0.0  
25:00 226.4 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 21:11  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

04/25/2022 12:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_