

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 002332  
11:19 29/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 2:00  
STERILIZING 3:06  
136.4 C 226kPa 3:06  
136.4 C 225kPa 3:36  
136.3 C 223kPa 4:06  
136.3 C 224kPa 4:36  
136.4 C 226kPa 5:06  
136.3 C 225kPa 5:36  
136.2 C 223kPa 6:06  
136.4 C 223kPa 6:36  
136.4 C 223kPa 6:36  
VENTING 6:36  
AIR DRYING 7:26  
CYCLE COMPLETE 9:37  
-----

Digital Signature #  
B4A86F60364F51A6  
-----