

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001558  
15:29 01/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:43  
STERILIZING 3:36  
136.0 C 220kPa 3:36  
135.6 C 221kPa 4:06  
136.2 C 223kPa 4:36  
136.4 C 225kPa 5:06  
136.3 C 225kPa 5:36  
136.4 C 225kPa 6:06  
136.4 C 225kPa 6:36  
136.1 C 222kPa 7:06  
136.1 C 222kPa 7:06  
VENTING 7:06  
AIR DRYING 7:56  
CYCLE COMPLETE 37:56  
-----

Digital Signature #  
943ADAB628EF9885  
-----