
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000556
15:35 23/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:34
STERILIZING 3:41
136.0 C 230kPa 3:41
136.4 C 233kPa 4:11
136.2 C 231kPa 4:41
136.1 C 231kPa 5:11
136.4 C 233kPa 5:41
136.3 C 233kPa 6:11
136.4 C 232kPa 6:41
136.4 C 231kPa 7:11
136.4 C 231kPa 7:11
VENTING 7:11
AIR DRYING 8:05
CYCLE COMPLETE 23:05

Digital Signature #
6F87900A628A8C0D
