
Midmark M11 - v1.0.0
Total Cycles: 3810

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 166.8 F 2.4
4:00 187.4 F 6.1
6:00 205.1 F 10.6
8:00 224.9 F 5.7
10:00 228.8 F 10.1
12:00 235.5 F 8.4
14:00 242.6 F 12.1
16:00 240.8 F 10.5
18:00 253.2 F 16.6
20:00 263.8 F 22.9
22:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.4
1:00 272.0 F 28.4
2:00 271.9 F 28.4
3:00 272.0 F 28.5
4:00 271.9 F 28.7

Min 271.5 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 246.5 F 0.0
10:00 228.5 F 0.0
15:00 190.7 F 0.0
20:00 197.6 F 0.0
25:00 172.4 F 0.0
Dry Cycle Aborted

FILLING: 1:16
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:37
DRYING: 28:20
TOTAL CYCLE: 00:59:50

09/25/2023 02:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____