
Midmark M11 - v1.0.5
Total Cycles: 910

Name: RITTER2
Sterilizer ID: V2313228
03/25/2022 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.0 F 0.0
2:00 172.8 F 1.5
4:00 201.4 F 6.6
6:00 228.1 F 12.0
8:00 228.4 F 8.3
10:00 238.2 F 11.7
12:00 242.3 F 12.6
14:00 242.7 F 12.2
16:00 255.8 F 19.1
18:00 267.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.4 F 29.4
2:00 272.8 F 29.8
3:00 272.5 F 30.1
4:00 271.6 F 29.3

Min 271.5 F 29.3
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 225.2 F 0.0
10:00 210.2 F 0.0
15:00 225.1 F 0.0
20:00 182.9 F 0.0
25:00 195.4 F 0.0
30:00 186.8 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

03/25/2022 03:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____