
Midmark M11 - v1.0.0
Total Cycles: 2121

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.8 F 0.0
2:00 159.5 F 0.5
4:00 176.2 F 3.7
6:00 208.8 F 7.6
8:00 209.3 F 12.6
10:00 225.2 F 6.1
12:00 236.7 F 10.8
14:00 231.3 F 7.1
16:00 244.1 F 12.9
18:00 241.6 F 11.1
20:00 253.8 F 17.4
22:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 29.0
2:00 271.3 F 28.2
3:00 272.5 F 29.1
4:00 271.6 F 28.4

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 223.9 F 0.0
10:00 237.0 F 0.0
15:00 224.0 F 0.0
20:00 228.8 F 0.0
25:00 231.3 F 0.0
30:00 234.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

07/13/2022 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____