
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001079
8:48 17/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:48
STERILIZING 4:04
136.5 C 224kPa 4:04
136.4 C 224kPa 4:34
136.3 C 224kPa 5:04
136.4 C 224kPa 5:34
136.2 C 223kPa 6:04
136.2 C 224kPa 6:34
136.4 C 225kPa 7:04
136.4 C 225kPa 7:34
136.4 C 225kPa 7:34
VENTING 7:34
AIR DRYING 8:24
CYCLE COMPLETE 23:24

Digital Signature #
2B8C099ECC0CAB59
