
Midmark M11 - v1.0.0
Total Cycles: 4851

Name: RITTER 1
Sterilizer ID: V2178198
08/26/2024 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.2 F 0.0
2:00 130.1 F 0.3
4:00 154.6 F 0.7
6:00 173.4 F 1.1
8:00 196.3 F 2.8
10:00 208.5 F 6.4
12:00 217.6 F 10.1
14:00 229.2 F 9.8
16:00 231.7 F 9.2
18:00 239.9 F 11.2
20:00 241.5 F 11.6
22:00 239.0 F 9.9
24:00 251.7 F 16.0
26:00 262.4 F 22.1
28:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.8
2:00 272.2 F 28.5
3:00 271.5 F 28.0
4:00 271.8 F 28.4

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 197.7 F 0.0
10:00 197.4 F 0.0
15:00 210.7 F 0.0
20:00 232.3 F 0.0
25:00 232.8 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 28:37
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:07:08
CYCLE COMPLETE

08/26/2024 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____