
Midmark M11 - v1.0.5
Total Cycles: 2900

Name: RITTER2
Sterilizer ID: V2313228
09/18/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 175.4 F 3.5
4:00 197.0 F 8.3
6:00 212.4 F 13.2
8:00 221.8 F 8.4
10:00 234.1 F 12.5
12:00 238.4 F 11.6
14:00 239.2 F 10.5
16:00 252.4 F 17.1
18:00 263.8 F 23.7
20:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.3 F 29.5
2:00 272.7 F 29.9
3:00 271.6 F 29.5
4:00 272.3 F 29.6

Min 271.3 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 227.6 F 0.0
10:00 220.1 F 0.0
15:00 211.0 F 0.0
20:00 179.8 F 0.0
25:00 175.7 F 0.0
30:00 173.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

09/18/2023 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____