
Midmark M11 - v1.0.0
Total Cycles: 2133

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 77.5 F 0.0
4:00 85.3 F 0.0
6:00 114.5 F 0.2
8:00 150.3 F 0.7
10:00 175.9 F 3.7
12:00 193.8 F 6.7
14:00 206.1 F 10.4
16:00 223.1 F 7.5
18:00 226.9 F 7.6
20:00 233.9 F 11.7
22:00 233.9 F 7.9
24:00 243.8 F 13.5
26:00 243.0 F 11.6
28:00 254.8 F 17.7
30:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.0 F 28.4
2:00 272.1 F 28.7
3:00 272.5 F 29.1
4:00 271.9 F 28.8

Min 271.5 F 28.3
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 200.6 F 0.0
10:00 207.0 F 0.0
15:00 212.1 F 0.0
20:00 215.5 F 0.0
25:00 241.0 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

07/15/2022 08:02 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____