

-----  
Midmark M11 - v1.0.5  
Total Cycles: 952

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2022 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 179.4 F 2.5  
4:00 202.0 F 8.2  
6:00 221.3 F 10.5  
8:00 233.7 F 9.9  
10:00 235.1 F 8.6  
12:00 241.8 F 11.2  
14:00 248.0 F 14.5  
16:00 261.3 F 21.9  
18:00 272.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.7 F 29.3  
2:00 271.3 F 28.8  
3:00 272.6 F 29.4  
4:00 271.7 F 28.7  
  
Min 271.2 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.0 F 0.0  
5:00 230.0 F 0.0  
10:00 228.1 F 0.0  
15:00 230.1 F 0.0  
20:00 229.5 F 0.0  
25:00 233.8 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:02  
CYCLE COMPLETE

04/04/2022 01:42 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02