
Midmark M11 - v1.0.0
Total Cycles: 750

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 193.3 F 3.3
4:00 194.6 F 8.2
6:00 209.5 F 9.6
8:00 223.8 F 9.6
10:00 234.7 F 7.7
12:00 243.8 F 12.8
14:00 242.4 F 11.3
16:00 256.0 F 18.3
18:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.2
2:00 272.6 F 28.8
3:00 271.7 F 28.8
4:00 272.1 F 29.3

Min 271.5 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 238.2 F 0.0
10:00 223.8 F 0.0
15:00 231.1 F 0.0
20:00 228.0 F 0.0
25:00 229.8 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:44
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

03/25/2021 03:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____