
Midmark M11 - v1.0.5
Total Cycles: 3121

Name: RITTER2
Sterilizer ID: V2313228
11/14/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 73.8 F 0.0
4:00 84.2 F 0.3
6:00 95.3 F 1.5
8:00 132.2 F 4.3
10:00 164.9 F 7.8
12:00 186.5 F 11.5
14:00 210.0 F 3.8
16:00 216.9 F 7.9
18:00 230.1 F 12.0
20:00 231.0 F 8.0
22:00 243.6 F 13.4
24:00 241.6 F 11.6
26:00 253.6 F 17.6
28:00 264.2 F 23.9
30:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 272.1 F 29.3
2:00 272.5 F 29.4
3:00 272.7 F 29.6
4:00 272.4 F 29.9

Min 271.9 F 29.3
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 205.5 F 0.0
10:00 197.8 F 0.0
15:00 186.8 F 0.0
20:00 185.3 F 0.0
25:00 183.8 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

11/14/2023 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____