

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3868

Name: RITTER2  
Sterilizer ID: V2313228  
08/16/2024 01:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.0 F 0.0  
2:00 90.1 F 0.0  
4:00 103.4 F 0.5  
6:00 138.0 F 2.7  
8:00 166.7 F 5.6  
10:00 195.0 F 8.8  
12:00 206.7 F 12.5  
14:00 214.0 F 5.7  
16:00 227.3 F 9.3  
18:00 236.5 F 13.4  
20:00 234.3 F 10.0  
22:00 237.9 F 9.6  
24:00 244.8 F 13.5  
26:00 255.5 F 19.2  
28:00 265.6 F 25.4  
30:00 271.2 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 271.6 F 29.8  
2:00 272.4 F 30.4  
3:00 272.7 F 30.3  
4:00 272.7 F 30.2

Min 271.0 F 29.2  
Max 272.8 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 203.5 F 0.0  
10:00 202.3 F 0.0  
15:00 195.0 F 0.0  
20:00 191.8 F 0.0  
25:00 194.0 F 0.0  
30:00 210.3 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 30:04  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:08:54  
CYCLE COMPLETE

08/16/2024 02:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_