

-----  
Midmark M11 - v1.0.5  
Total Cycles: 779

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2022 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
2:00 178.3 F 0.0  
4:00 199.4 F 1.9  
6:00 215.9 F 5.6  
8:00 230.8 F 10.4  
10:00 234.2 F 8.4  
12:00 243.6 F 12.2  
14:00 245.5 F 13.3  
16:00 245.6 F 13.3  
18:00 258.9 F 20.5  
20:00 270.5 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.7 F 29.2  
2:00 271.8 F 28.8  
3:00 272.7 F 29.2  
4:00 271.7 F 28.5

Min 271.6 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 230.9 F 0.0  
10:00 226.6 F 0.0  
15:00 215.6 F 0.0  
20:00 212.9 F 0.0  
25:00 216.5 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 20:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

02/10/2022 11:09 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02