

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4945

Name: RITTER 1  
Sterilizer ID: V2178198  
10/01/2024 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 76.9 F 0.0  
4:00 117.1 F 0.1  
6:00 116.1 F 0.3  
8:00 142.7 F 0.4  
10:00 169.5 F 0.9  
12:00 194.9 F 4.1  
14:00 207.3 F 7.4  
16:00 218.2 F 10.9  
18:00 223.3 F 5.9  
20:00 232.2 F 9.0  
22:00 242.0 F 12.6  
24:00 240.7 F 10.9  
26:00 236.7 F 8.9  
28:00 250.8 F 15.4  
30:00 262.0 F 21.8  
32:00 272.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.6  
2:00 271.5 F 27.9  
3:00 272.7 F 28.7  
4:00 271.7 F 27.9  
  
Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 194.7 F 0.0  
10:00 206.7 F 0.0  
15:00 214.8 F 0.0  
20:00 217.7 F 0.0  
25:00 220.1 F 0.0  
30:00 219.1 F 0.0  
  
Drying Complete

FILLING: 0:43  
HEATING: 32:30  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:10:35  
CYCLE COMPLETE

10/01/2024 09:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_