

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2914

Name: RITTER 1  
Sterilizer ID: V2178198  
02/16/2023 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
2:00 188.6 F 3.8  
4:00 195.0 F 8.4  
6:00 215.2 F 11.3  
8:00 227.9 F 8.9  
10:00 236.2 F 10.0  
12:00 240.7 F 11.7  
14:00 239.2 F 10.0  
16:00 252.7 F 16.7  
18:00 264.2 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.2 F 28.9  
2:00 271.4 F 28.4  
3:00 272.2 F 28.7  
4:00 271.5 F 28.6  
  
Min 271.3 F 28.3  
Max 272.3 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 234.8 F 0.0  
10:00 229.0 F 0.0  
15:00 230.6 F 0.0  
20:00 228.4 F 0.0  
25:00 221.5 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 19:47  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

02/16/2023 12:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_