

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4545

Name: RITTER 1  
Sterilizer ID: V2178198  
05/13/2024 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 74.8 F 0.0  
4:00 80.1 F 0.2  
6:00 111.0 F 1.6  
8:00 136.4 F 3.9  
10:00 160.8 F 6.6  
12:00 185.1 F 9.4  
14:00 202.7 F 12.9  
16:00 211.0 F 4.3  
18:00 221.2 F 7.6  
20:00 230.1 F 11.2  
22:00 228.7 F 6.6  
24:00 238.7 F 11.1  
26:00 234.2 F 8.0  
28:00 247.3 F 14.0  
30:00 257.5 F 19.5  
32:00 266.6 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.1  
2:00 271.8 F 28.4  
3:00 271.9 F 28.9  
4:00 272.4 F 28.9  
  
Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 199.4 F 0.0  
10:00 185.3 F 0.0  
15:00 200.2 F 0.0  
20:00 218.4 F 0.0  
25:00 232.7 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:51  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:12:07  
CYCLE COMPLETE

05/13/2024 09:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_