

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 008446  
8:22 25/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 3:54  
STERILIZING 5:07  
136.0 C 220kPa 5:07  
136.3 C 221kPa 5:37  
136.1 C 217kPa 6:07  
136.4 C 220kPa 6:37  
136.2 C 220kPa 7:07  
135.9 C 220kPa 7:37  
136.3 C 222kPa 8:07  
136.3 C 220kPa 8:37  
136.3 C 220kPa 8:37  
VENTING 8:37  
AIR DRYING 9:28  
CYCLE COMPLETE 10:22

-----  
Digital Signature #  
F570FCD03F227428  
-----