
Midmark M11 - v1.0.0
Total Cycles: 718

Name: MIDMARK
Sterilizer ID: V2178198
03/02/2021 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.8 F 0.0
2:00 127.9 F 0.8
4:00 153.0 F 3.6
6:00 179.0 F 7.9
8:00 202.4 F 12.9
10:00 215.6 F 6.1
12:00 230.2 F 11.3
14:00 234.1 F 8.1
16:00 240.6 F 10.7
18:00 245.9 F 13.0
20:00 258.9 F 20.0
22:00 270.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.2 F 29.0
2:00 272.7 F 29.2
3:00 272.8 F 29.2
4:00 272.8 F 29.2

Min 271.8 F 28.6
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 209.8 F 0.0
10:00 208.1 F 0.0
15:00 206.7 F 0.0
20:00 203.6 F 0.0
25:00 203.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:25
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

03/02/2021 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____