
Midmark M11 - v1.0.0
Total Cycles: 816

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.9 F 0.0
2:00 165.3 F 2.8
4:00 178.6 F 6.0
6:00 200.8 F 10.8
8:00 214.5 F 4.4
10:00 228.0 F 9.0
12:00 237.0 F 11.8
14:00 239.6 F 11.1
16:00 237.8 F 9.4
18:00 251.3 F 15.9
20:00 262.8 F 22.5
22:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.9 F 29.0
2:00 271.7 F 28.2
3:00 272.8 F 29.1
4:00 272.3 F 28.6

Min 271.6 F 28.2
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 213.7 F 0.0
10:00 214.5 F 0.0
15:00 202.7 F 0.0
20:00 201.9 F 0.0
25:00 213.7 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

04/27/2021 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____