
Midmark M11 - v1.0.0
Total Cycles: 2751

Name: RITTER 1
Sterilizer ID: V2178198
01/09/2023 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 160.3 F 0.0
4:00 180.8 F 1.3
6:00 208.6 F 5.4
8:00 214.9 F 10.6
10:00 226.7 F 6.0
12:00 235.6 F 11.0
14:00 233.9 F 8.2
16:00 240.8 F 10.8
18:00 246.6 F 13.4
20:00 259.2 F 20.3
22:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.8 F 28.4
3:00 271.2 F 28.4
4:00 272.2 F 29.0

Min 271.0 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 214.3 F 0.0
10:00 222.9 F 0.0
15:00 224.3 F 0.0
20:00 234.6 F 0.0
25:00 229.9 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

01/09/2023 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____